



No.	Clarification requested	Answer
1	The PSA plant required, is it Duplex ( 2x20Nm <sup>3</sup> /hr=40Nm <sup>3</sup> /hr or 2x40Nm <sup>3</sup> /hr=80Nm <sup>3</sup> /hr) or a Single PSA plant(40Nm <sup>3</sup> /hr with quantity 2)?	There are two separate facilities that we are procuring for, each of which will receive a plant solution. This two plant solutions, each with a total rated output of up to 40 Nm <sup>3</sup> /hr. Duplexing of any of the components (especially compressors) is preferred; however, due to space limitations for a containerized solution, duplexing is not a requirement for a successful technical offer. Manufacturers/vendors to propose appropriate and feasible solution.
2	In Annex B, the configuration of the PSA plant stated is a Duplex system. But in Annex A, two custom containers are requested. Kindly clarify whether all the components should be installed in one container or treat each PSA Plant and accessory as separate entities to be placed in separate containers.	The preference is for each site to be one (1) container for housing with PSA solution contained therein. With regard to cylinders, regulators, flowmeters, humidifiers, we are not seeking housing/storage, just transport.
3	Can we submit more than one model for the cylinder accessories?	As long as quality documentation packages are clear, valid, and complete, we are willing to review multiple options proposed by a single vendor.
4	I would like confirmation on your payment terms. Since COVID, most medical (oxygen) device manufactures /suppliers request upfront payment before shipping products. Due to this scenario, we would like to clarify if it will be possible to negotiate an upfront payment of 50% followed by a 50% payment after supply and inspection?	We are open to payment negotiations, to be discussed during contracting phase. Any requests as such can be indicated on the offer paperwork.
5	Kindly confirm that the consignee will be Clinton Health Access Initiative - DRC.	Details regarding the consignee will be shared during contracting. The proposal shall be directed to the Clinton Health Access Initiative.
6	We would like to request if you can send some clarifications request in order to present our proposal on 20th May on time	Our interpretation of this ask is that we share all questions and answers with you prior to the submission deadline. As per our processes, we share all questions and answers with all parties who have inquired, and we post them publicly where the RFP has been floated: <a href="https://www.clintonhealthaccess.org/request-for-proposals-rfp-for-containerized-psa-oxygen-plant-in-drc/">https://www.clintonhealthaccess.org/request-for-proposals-rfp-for-containerized-psa-oxygen-plant-in-drc/</a>